

各確診檢驗單位 FT₄ 院際品管定量檢驗之再現性

送檢批次：CHT2011-04、CHT2011-11、CHT2012-02

	Method code	n	Mean	SD	CV
RH01b	02	3	1.47	0.08	5.4%
RH02a	02	3	1.55	0.21	13.5%
RH07b	02	3	1.62	0.19	11.7%
RH14	02	3	1.09	0.01	0.9%
CL006b	02	3	1.56	0.27	17.3%
CL009	02	3	1.06	0.11	10.4%
CL010	02	3	1.09	0.05	4.6%
CL012	02	3	0.99	0.07	7.1%
CL013	02	3	1.64	0.08	4.9%
CL015	02	3	2.04	0.07	3.4%
Median for Mean, SD and CV			1.51	0.08	6.3%
Range for Mean, SD and CV			0.99 ~ 2.04	0.01 ~ 0.27	0.9% ~ 17.3%
RH01a	04	3	1.35	0.06	4.4%
RH02c	04	3	1.37	0.04	2.9%
RH06	04	3	1.45	0.07	4.8%
RH07a	04	3	1.32	0.09	6.8%
RH12	04	3	1.30	0.09	6.9%
RH19	04	3	1.23	0.06	4.9%
RH20	04	3	1.55	0.12	7.7%
CL005	04	3	1.40	0.04	2.9%
CL006a	04	3	1.44	0.08	5.6%
CL008	04	3	1.57	0.07	4.5%
CL011	04	3	1.33	0.08	6.0%
CL014b	04	3	1.57	0.03	1.9%
Median for Mean, SD and CV			1.37	0.07	4.9%
Range for Mean, SD and CV			1.23 ~ 1.57	0.04~0.12	1.9% ~ 7.7%
Median of All methods median			1.42		0.08
Range of All methods Range			0.99 ~ 2.04		0.01 ~ 0.27

Method code 02 : RIA (Competitive-type) ; Method code 04 : CLIA (Competitive-type)