

各確診檢驗單位 T₄ 院際品管定量檢驗之再現性

送檢批次：CHT2011-04、CHT2011-11、CHT2012-02

	Method code	n	Mean	SD	CV
RH01b	02	3	6.6	0.46	7.0%
RH02a	02	3	5.8	1.10	19.0%
RH07b	02	3	6.7	0.96	14.3%
RH14	02	3	6.7	0.00	0.0%
RH15	02	3	6.9	0.61	8.8%
CL006b	02	3	6.7	0.25	3.7%
CL009	02	3	6.5	0.38	5.8%
CL010	02	3	6.3	0.21	3.3%
CL012	02	3	6.8	0.30	4.4%
CL014a	02	3	6.5	1.78	27.4%
CL015	02	3	6.4	0.42	6.6%
Median for Mean, SD and CV			6.6	0.42	6.6%
Range for Mean, SD and CV			5.8 ~ 6.9	0.00 ~ 1.78	0.0% ~ 27.4%
RH01a	04	3	8.1	0.36	4.4%
RH02c	04	3	7.7	0.86	11.2%
RH06	04	3	7.7	0.40	5.2%
RH07a	04	3	7.2	0.25	3.5%
RH12	04	3	7.1	0.51	7.2%
RH19	04	3	7.1	0.67	9.4%
RH20	04	3	7.2	0.30	4.2%
CL005	04	3	7.4	0.75	10.1%
CL006a	04	3	7.5	0.61	8.1%
CL008	04	3	6.7	0.26	3.9%
CL011	04	3	7.7	0.70	9.1%
CL013	04	3	0.36	0.36	4.4%
Median for Mean, SD and CV			7.4	0.51	7.2%
Range for Mean, SD and CV			6.7 ~ 8.1	0.25 ~ 0.86	3.5% ~ 11.2%
Median of All methods median			6.8	0.46	7.0%
Range of All methods Range			5.3 ~ 8.1	0.00 ~ 1.78	0.0% ~ 27.4%

Method code 02 : RIA (Competitive-type) ; Method code 04 : CLIA (Competitive-type)