

各確診檢驗單位 TSH 院際品管定量檢驗之再現性

送檢批次：CHT2011-04、CHT2011-11、CHT2012-02

	Method code	n	Mean	SD	CV
RH01b	01	3	5.6	0.29	5.2%
RH02a	01	3	5.3	0.15	2.8%
RH07b	01	3	5.5	0.26	4.7%
RH14	01	3	4.2	0.25	6.0%
RH15	01	3	5.2	0.40	7.7%
CL006b	01	3	5.5	0.32	5.8%
CL009	01	3	5.5	0.10	1.8%
CL010	01	3	5.6	0.21	3.8%
CL012	01	3	5.6	0.20	3.6%
CL013	01	3	5.3	0.52	9.8%
CL014a	01	3	5.1	0.17	3.3%
CL015	01	3	5.5	0.25	4.5%
Median for Mean, SD and CV			5.5	0.25	4.6%
Range for Mean, SD and CV			4.2 ~ 5.6	0.10 ~ 0.52	1.8% ~ 9.8%
RH01a	03	3	3.8	0.25	6.6%
RH02c	03	3	3.8	0.30	7.9%
RH06	03	3	3.9	0.23	5.9%
RH07a	03	3	4.0	0.00	0.0%
RH12	03	3	4.1	0.17	4.1%
RH19	03	3	4.1	0.17	4.1%
RH20	03	3	4.2	0.21	5.0%
CL005	03	3	3.9	0.20	5.1%
CL006a	03	3	3.8	0.25	6.6%
CL008	03	3	3.6	0.06	1.7%
CL011	03	3	3.9	0.25	6.4%
Median for Mean, SD and CV			3.9	0.21	5.1%
Range for Mean, SD and CV			3.6 ~ 4.2	0.00 ~ 0.30	0.0% ~ 7.9%
Median of All methods median			4.2	0.23	5.0%
Range of All methods Range			3.6 ~ 5.6	0.00 ~ 0.52	0.0% ~ 9.8%

Method code 01 : RIA (Sandwich-type) ; Method code 03 : CLIA (Sandwich-type)