

各確診檢驗單位 FT₄ 院際品管定量檢驗之中間精密度 (S2)

送檢批次：CHT2012-04-S2、CHT2012-08-S1、CHT2013-04-S2

	Method code	n	Mean	SD	CV
RH01b	02	3	1.57	0.07	4.5%
RH07b	02	3	2.00	0.22	11.0%
RH14	02	3	1.25	0.14	11.2%
CL009 ^{註1}	02	3	1.34	0.29	21.6%
CL010	02	3	1.25	0.08	6.4%
CL012 ^{註1}	02	3	1.46	0.43	29.5%
CL013	02	3	1.69	0.13	7.7%
CL015	02	3	2.11	0.18	8.5%
Median for Mean, SD and CV			1.52	0.16	9.8%
Range for Mean, SD and CV			1.25 ~ 2.11	0.07 ~ 0.43	4.5% ~ 29.5%
RH01a	04	3	2.22	0.17	7.7%
RH02c ^{註1}	04	3	2.03	0.48	23.6%
RH06	04	3	2.23	0.16	7.2%
RH07a	04	3	2.13	0.29	13.6%
RH12	04	3	2.06	0.06	2.9%
RH19	04	3	1.96	0.13	6.6%
RH20	04	3	1.81	0.10	5.5%
CL005	04	3	2.29	0.08	3.5%
CL006a	04	3	1.78	0.04	2.2%
CL008	04	3	1.61	0.02	1.2%
CL011 ^{註2}	04	3	2.23	0.09	4.0%
CL014b	04	3	1.54	0.03	1.9%
Median for Mean, SD and CV			2.05	0.10	4.8%
Range for Mean, SD and CV			1.54 ~ 2.29	0.02 ~ 0.48	1.2% ~ 23.6%
Median for All methods			1.89	0.13	6.9%
Range for All methods			1.25 ~ 2.29	0.02 ~ 0.48	1.2% ~ 29.5%

Method code 02 : RIA (Competitive-type) ; Method code 04 : CLIA (Competitive-type)

註 1 : 該單位於 CHT2013-04 更換試藥廠牌

註 2 : 該單位自 CHT2012-08 更換試藥廠牌