

各確診檢驗單位 T<sub>3</sub> 院際品管定量檢驗之中間精密度 (S2)

送檢批次：CHT2012-04-S2、CHT2012-08-S1、CHT2013-04-S2

	Method code	n	Mean	SD	CV
RH01b	02	3	144	5.9	4.1%
RH07b	02	3	108	6.0	5.6%
RH14 <sup>註1</sup>	02	3	118	6.8	5.8%
RH15	02	3	123	10.3	8.4%
CL009	02	3	133	10.1	7.6%
CL010	02	3	126	8.4	6.7%
CL012	02	3	114	9.8	8.6%
CL013	02	3	137	3.8	2.8%
CL014a	02	3	127	15.5	12.2%
CL015	02	3	131	4.0	3.1%
Median for Mean, SD and CV			127	7.6	6.3%
Range for Mean, SD and CV			108 ~ 144	3.8 ~ 15.5	2.8% ~ 12.2%
RH01a	04	3	177	13.1	7.4%
RH02c <sup>註1</sup>	04	3	168	14.7	8.8%
RH06	04	3	174	1.5	0.9%
RH07a	04	3	143	2.1	1.5%
RH12	04	3	138	10.4	7.5%
RH19	04	3	140	4.4	3.1%
RH20	04	3	180	19.6	10.9%
CL005	04	3	177	4.6	2.6%
CL006a	04	3	141	1.7	1.2%
CL008	04	3	149	2.1	1.4%
CL011 <sup>註2</sup>	04	3	188	19.3	10.3%
Median for Mean, SD and CV			168	4.6	3.1%
Range for Mean, SD and CV			138 ~ 188	1.5 ~ 19.6	0.9% ~ 10.9%
Median for All methods			140	6.8	5.8%
Range for All methods			108 ~ 188	1.5 ~ 19.6	0.9% ~ 12.2%

Method code 02 : RIA ( Competitive-type ) ; Method code 04 : CLIA ( Competitive-type )

註 1 : 該單位於 CHT2013-04 更換試藥廠牌

註 2 : 該單位自 CHT2012-08 更換試藥廠牌