

各確診檢驗單位 T₄ 院際品管定量檢驗之中間精密度 (S2)

送檢批次：CHT2012-04-S2、CHT2012-08-S1、CHT2013-04-S2

	Method code	n	Mean	SD	CV
RH01b	02	3	7.2	0.06	0.8%
RH07b	02	3	8.2	0.31	3.8%
RH15	02	3	8.0	0.32	4.0%
CL009	02	3	7.4	0.40	5.4%
CL010	02	3	7.6	0.31	4.1%
CL012 ^{註1}	02	3	7.7	0.50	6.5%
CL014a	02	3	7.3	0.26	3.6%
CL015	02	3	7.7	1.15	14.9%
Median for Mean, SD and CV			7.7	0.32	4.1%
Range for Mean, SD and CV			7.2 ~ 8.2	0.06 ~ 1.15	0.8% ~ 14.9%
RH01a	04	3	8.7	0.36	4.1%
RH02c ^{註1}	04	3	8.3	0.85	10.2%
RH06	04	3	8.7	0.66	7.6%
RH07a	04	3	9.7	0.66	6.8%
RH12	04	3	9.2	0.45	4.9%
RH14 ^{註2}	04	1	8.6	-	-
RH19	04	3	9.1	0.47	5.2%
RH20	04	3	8.7	0.21	2.4%
CL005	04	3	8.5	0.80	9.4%
CL006a	04	3	8.9	0.26	2.9%
CL008	04	3	7.7	0.32	4.2%
CL011 ^{註3}	04	3	8.5	0.60	7.1%
CL013 ^{註3}	04	3	8.4	1.47	17.5%
Median for Mean, SD and CV			8.7	0.54	6.0%
Range for Mean, SD and CV			7.7 ~ 9.7	0.21 ~ 1.47	2.4% ~ 17.5%
Median for All methods			8.4	0.43	5.1%
Range for All methods			7.2 ~ 9.7	0.06 ~ 1.47	0.8% ~ 17.5%

Method code 02 : RIA (Competitive-type) ; Method code 04 : CLIA (Competitive-type)

註 1 : 該單位於 CHT2013-04 更換試藥廠牌

註 2 : 該單位於 CHT2013-04 更換檢驗方法

註 3 : 該單位自 CHT2012-08 更換試藥廠牌