

各確診檢驗單位 TSH 院際品管定量檢驗之中間精密度 (S2)

送檢批次：CHT2012-04-S2、CHT2012-08-S1、CHT2013-04-S2

	Method code	n	Mean	SD	CV
RH01b	01	3	14.9	0.89	6.0%
RH07b	01	3	14.9	0.74	5.0%
RH14 ^{註1}	01	3	18.0	2.16	12.0%
RH15	01	3	15.4	0.21	1.4%
CL009	01	3	14.8	0.56	3.8%
CL010	01	3	15.0	0.12	0.8%
CL012	01	3	15.8	0.74	4.7%
CL013	01	3	14.8	2.86	19.3%
CL014a	01	3	15.0	1.00	6.7%
CL015	01	3	14.8	0.42	2.8%
Median for Mean, SD and CV			15.0	0.74	4.9%
Range for Mean, SD and CV			14.8 ~ 18.0	0.12 ~ 2.86	0.8% ~ 19.3%
RH01a	03	3	13.7	0.72	5.3%
RH02c ^{註1}	03	3	13.9	1.62	11.7%
RH06	03	3	14.6	1.95	13.4%
RH07a	03	3	12.6	0.25	2.0%
RH12	03	3	14.6	0.76	5.2%
RH19	03	3	15.2	0.21	1.4%
RH20	03	3	15.0	0.55	3.7%
CL005	03	3	13.2	0.20	1.5%
CL006a	03	3	15.0	0.10	0.7%
CL008	03	3	12.4	0.36	2.9%
CL011 ^{註2}	03	3	13.6	0.95	7.0%
Median for Mean, SD and CV			13.9	0.55	3.7%
Range for Mean, SD and CV			12.4 ~ 15.2	0.10 ~ 1.95	0.7% ~ 13.4%
Median for All methods			14.8	0.72	4.7%
Range for All methods			12.4 ~ 18.0	0.10 ~ 2.86	0.7% ~ 19.3%

Method code 01 : RIA (Sandwich-type) ; Method code 03 : CLIA (Sandwich-type)

註 1 : 該單位於 CHT2013-04 更換試藥廠牌

註 2 : 該單位自 CHT2012-08 更換試藥廠牌