

各確診檢驗單位 FT₄ 院際品管定量檢驗之中間精密度 (S3)

送檢批次：CHT2012-04-S3、CHT2012-06-S2、CHT2013-04-S3

	Method code	n	Mean	SD	CV
RH01b	02	3	2.19	0.08	3.7%
RH07b	02	3	3.20	0.41	12.8%
RH14	02	3	1.75	0.17	9.7%
CL009 ^{註1}	02	3	1.90	0.26	13.7%
CL010	02	3	1.71	0.06	3.5%
CL012 ^{註1}	02	3	2.01	0.59	29.4%
CL013	02	3	2.31	0.09	3.9%
CL015	02	3	2.92	0.31	10.6%
Median for Mean, SD and CV			2.10	0.22	10.2%
Range for Mean, SD and CV			1.71 ~ 3.20	0.06 ~ 0.59	3.5% ~ 29.4%
RH01a	04	3	3.37	0.22	6.5%
RH02c ^{註1}	04	3	2.80	0.63	22.5%
RH06	04	3	3.17	0.18	5.7%
RH07a	04	3	3.00	0.23	7.7%
RH12	04	3	2.99	0.02	0.7%
RH19	04	3	3.18	0.07	2.2%
RH20	04	3	2.57	0.03	1.2%
CL005	04	3	3.38	0.09	2.7%
CL006a	04	3	2.56	0.07	2.7%
CL008	04	3	2.31	0.07	3.0%
CL011 ^{註2}	04	3	3.18	0.16	5.0%
CL014b	04	3	1.87	0.61	32.6%
Median for Mean, SD and CV			3.00	0.13	4.0%
Range for Mean, SD and CV			1.87 ~ 3.38	0.02 ~ 0.63	0.7% ~ 32.6%
Median for All methods			2.69	0.17	5.4%
Range for All methods			1.71 ~ 3.38	0.02 ~ 0.63	0.7% ~ 32.6%

Method code 02 : RIA (Competitive-type) ; Method code 04 : CLIA (Competitive-type)

註 1 : 該單位於 CHT2013-04 更換試藥廠牌

註 2 : 該單位自 CHT2012-08 更換試藥廠牌