

各確診檢驗單位 T₃ 院際品管定量檢驗之中間精密度 (S3)

送檢批次：CHT2012-04-S3、CHT2012-06-S2、CHT2013-04-S3

	Method code	n	Mean	SD	CV
RH01b	02	3	148	5.9	4.0%
RH07b	02	3	126	9.0	7.1%
RH14 ^{註1}	02	3	126	11.1	8.8%
RH15	02	3	127	5.5	4.3%
CL009	02	3	141	3.5	2.5%
CL010	02	3	137	9.1	6.6%
CL012	02	3	131	10.0	7.6%
CL013	02	3	156	11.5	7.4%
CL014a	02	3	131	6.4	4.9%
CL015	02	3	142	17.1	12.0%
Median for Mean, SD and CV			134	9.1	6.9%
Range for Mean, SD and CV			126 ~ 156	3.5 ~ 17.1	2.5% ~ 12.0%
RH01a	04	3	197	14.2	7.2%
RH02c ^{註1}	04	3	189	24.6	13.0%
RH06	04	3	179	9.1	5.1%
RH07a	04	3	150	2.6	1.7%
RH12	04	3	152	9.3	6.1%
RH19	04	3	156	6.5	4.2%
RH20	04	3	195	12.5	6.4%
CL005	04	3	188	7.0	3.7%
CL006a	04	3	157	4.0	2.5%
CL008	04	3	162	1.5	0.9%
CL011 ^{註2}	04	3	189	26.4	14.0%
Median for Mean, SD and CV			179	9.1	5.1%
Range for Mean, SD and CV			150 ~ 197	1.5 ~ 26.4	0.9% ~ 14.0%
Median for All methods			152	9.1	6.1%
Range for All methods			126 ~ 197	1.5 ~ 26.4	0.9% ~ 14.0%

Method code 02 : RIA (Competitive-type) ; Method code 04 : CLIA (Competitive-type)

註 1 : 該單位於 CHT2013-04 更換試藥廠牌

註 2 : 該單位自 CHT2012-08 更換試藥廠牌