

各確診檢驗單位 T₄ 院際品管定量檢驗之中間精密度 (S3)

送檢批次：CHT2012-04-S3、CHT2012-06-S2、CHT2013-04-S3

	Method code	n	Mean	SD	CV
RH01b	02	3	10.0	0.49	4.9%
RH07b	02	3	10.7	1.10	10.3%
RH15	02	3	10.8	0.55	5.1%
CL009	02	3	9.7	0.26	2.7%
CL010	02	3	10.1	0.55	5.4%
CL012 ^{註1}	02	3	9.5	0.44	4.6%
CL014a	02	3	9.9	0.21	2.1%
CL015	02	3	10.1	0.76	7.5%
Median for Mean, SD and CV			10.1	0.52	5.0%
Range for Mean, SD and CV			9.5 ~ 10.8	0.21 ~ 1.10	2.1% ~ 10.3%
RH01a	04	3	11.9	0.06	0.5%
RH02c ^{註1}	04	3	11.1	0.47	4.2%
RH06	04	3	11.4	0.32	2.8%
RH07a	04	3	12.8	0.58	4.5%
RH12	04	3	13.5	0.66	4.9%
RH14 ^{註2}	04	1	12.0	-	-
RH19	04	3	12.1	0.40	3.3%
RH20	04	3	11.1	0.00	0.0%
CL005	04	3	11.8	0.93	7.9%
CL006a	04	3	11.4	0.49	4.3%
CL008	04	3	11.0	0.15	1.4%
CL011 ^{註3}	04	3	11.5	1.04	9.0%
CL013 ^{註3}	04	3	13.3	2.26	17.0%
Median for Mean, SD and CV			11.8	0.48	4.3%
Range for Mean, SD and CV			11.0 ~ 13.5	0.00 ~ 2.26	0.0% ~ 17.0%
Median for All methods			11.1	0.49	4.6%
Range for All methods			9.5 ~ 13.5	0.00 ~ 2.26	0.0% ~ 17.0%

Method code 02 : RIA (Competitive-type) ; Method code 04 : CLIA (Competitive-type)

註 1 : 該單位於 CHT2013-04 更換試藥廠牌

註 2 : 該單位於 CHT2013-04 更換檢驗方法

註 3 : 該單位自 CHT2012-08 更換試藥廠牌